

Advanced Photon Source

User Policies and Procedures

POLICY	Page 1 of 3
Policy & Procedure #:	3.1.03
Revision #:	1
Issue Date:	8/26/10
Review Period:	1 year
Supersedes:	Rev. 0, 7/1/05
Last Reviewed:	8/16/11

Working on Beamline and Front-End Shielding Components

Changes made in this revision:

- Format and editorial changes including extraction from the integrated APS User Safety Policies and Procedures (3.1.102, APS document number APS_1258430) to make a stand-alone policy and procedure.
- Reference to specific legacy ESH Manual chapters deleted.

Prepared by:

Steve Davey, AES

Reviewed by:

AES/ESH Coordinator
AES/UES Group Leader
XSD/ESH Coordinator
AES/MIS ADD

Approved by:

AES/Division Director
APS/Deputy Director, X-ray Science
APS/Director

APS_1181699

The current version of this procedure is accessible from <http://centraldocs.aps.anl.gov/>. Print or electronically downloaded copies may be obsolete. Before using such a copy for work direction, employees must verify that it is current by comparing its revision number to that shown in the online version.

Policy

In recognition of the potentially significant hazards associated with beamline or front end radiation shielding, it is the policy of the APS that the APS Engineering Support (AES) Division is responsible for all work on beamline and front end shielding. This responsibility encompasses the labor for alignment, shielding validation, maintenance, repair, and any modification of any beamline shielding component.

If the beamline or front end component has an APS Radiation Shielding red tag, then the AES Division is responsible for any work on the component.

Users and other beamline personnel may operate engineered radiation safety shielding components. For example: users may actuate beamline shutters, and beamline personnel may open or close permanently installed manual shutters or stops, or may open or close an enclosure labyrinth or station door. The shielding systems must be used only as they were designed/intended and with APS and beamline prescribed safeguards (e.g., PSS safety interlocks and administrative controls).

There may be specific cases where the beamline management may seek to have beamline personnel assist with work on radiation shielding. Requests for this permission may be made to the AES Division Director. Beamline personnel may only work on the shielding with the written authorization of the AES Division Director or designee for a specific scope of work.

Labor

The AES Division will provide labor for the alignment, shielding validation, maintenance, and repair of beamline shielding components at no cost to the beamline management. AES, on a case-by-case basis, may recover labor costs from the CAT/CDT/XSD.

Parts

The beamline management provides parts (or the funding for the parts) for maintaining the beamline shielding components. Due to the diversity of parts required, often unique to one component on one beamline, and the beamline management's ownership of the hardware, the beamline management is responsible for maintaining spares. AES may maintain some stock of common spare parts (e.g., pneumatic actuator seals).

AES Technical Support

The AES Division personnel shall ensure the information (specifications, drawings, procedures, etc.) is available and provide the trained personnel to ensure that the work is done properly and the shielding device is returned to service safely.

Safety

In order to maintain a high level of safety, each person and group that requests or takes part in work on beamline shielding components should be aware that there are safety implications of their requests or activities. Further, if they are aware of a potentially unsafe situation resulting from the work, then the component or system is to be secured until the questions are resolved.

No work by any party on any radiation safety shielding is allowed without a complete, approved Configuration Control Work Permit (CCWP).¹ No work shall start on the beamline or front end shielding without the CCWP specified authorizations.

If the component is a new installation or a modification of an existing design, which operates outside of the safety/performance envelope of previous design reviews, the component shall not be brought into operation until the design and use are reviewed and approved by the appropriate design review committee.

For newly installed components, the Critical Component System Manager will work in concert with the Beamline Commissioning Readiness Review Team to ensure that the design/operational requirements records are in place.

References

¹[Configuration Control Work Permit Procedure \(APS 1192930\), # 3.1.06.](#)

FEEDBACK AND IMPROVEMENT

If you are using this procedure and have comments or suggested improvements for it, please go to the [APS Policies and Procedures Comment Form](#)^{*} to submit your input to a Procedure Administrator. If you are reviewing this procedure in workflow, your input must be entered in the comment box when you approve or reject the procedure.

Instructions for execution-time modifications to a policy/procedure can be found in the following document: Field Modification of APS Policy/Procedure ([APS 1408152](#)).

^{*} http://centraldocs.aps.anl.gov/comment_form.php