

Advanced Photon Source

PROCEDURE	Page 1 of 5
Procedure #:	3.1.52
Revision #:	0
Issue Date:	10/31/12
Review Period:	1 year
Supersedes:	N/A
Last Reviewed:	10/31/12

APS Protocol for Reporting Construction Vibration Issues for Beamlines

Section where used:

This procedure describes the process for Beamline Staff reporting suspected construction-related vibration issues to the APS Construction Vibration Measurement Task Force. It includes the steps to be used in reporting to the APS ALD Office and determination if construction activities need to be suspended.

Changes made in this revision:

- New procedure

Prepared by:

Bruce Glagola, APS User Safety Officer/UES, Construction Vibration Measurement Task Force Chair

Reviewed by:

XSD/Deputy Division Director

Julie Cross, AES Technical Support Services Group

Approved by:

AES/Division Director

APS_1432154

The current version of this procedure is accessible from <http://centraldocs.aps.anl.gov/>. Print or electronically downloaded copies may be obsolete. Before using such a copy for work direction, employees must verify that it is current by comparing its revision number to that shown in the online version.

Advanced Photon Source

PROCEDURE	Page 2 of 5
Procedure #:	3.1.52
Revision #:	0

Table of Contents

1	INTRODUCTION.....	3
1.1	Purpose	3
1.2	Scope	3
1.3	Applicability.....	3
1.4	Type of Procedure	3
1.5	Hazard Controls.....	3
2	BACKGROUND.....	3
3	PROCEDURE	4
4	FEEDBACK AND IMPROVEMENT	5

PROCEDURE	Page 3 of 5
Procedure #:	3.1.52
Revision #:	0

APS Protocol for Reporting Construction Vibration Issues for Beamlines

1 INTRODUCTION

1.1 Purpose

This procedure is used to define the steps to be taken when a beamline staff member reports that vibrations related to construction activities at the APS may be impacting beamline operation and data quality. The procedure includes guidance on the steps to be taken by beamline staff, Floor Coordinators, members of the APS Vibration Measurement Task Force Subteam 4, the APS ALD Office, MCR personnel, and construction managers. The decision to suspend construction activities will be made at the APS ALD level.

1.2 Scope

- Vibrations due to construction activities at the APS
- Suspension of construction activities if needed

1.3 Applicability

This procedure applies to vibrations caused by construction activity at the APS only.

1.4 Type of Procedure

This is a guidance procedure.

1.5 Hazard Controls

None required.

2 BACKGROUND

The APS has formed a Construction Vibration Measurement Task Force (CVMTF) to deploy vibration sensors (accelerometers) on the APS experiment hall floor. The CVMTF contains four subteams responsible for the following areas:

- Subteam 1 – Coordinates construction vibration equipment tests and on-site schedules of construction details.
- Subteam 2 – Ensures readiness of beamline accelerometers to collect data and to connect with EPICS as appropriate when construction equipment is being tested and used.

- Subteam 3 – Ensures the MCR is ready to monitor SR electron movement while construction equipment is being tested and used. As necessary, this subteam will communicate when electron beam motion is out of normal.
- Subteam 4 – Works with designated beamlines when construction vibration measurements are to be monitored. This subteam will interface with beamline staff if disruptive construction-related vibrations are reported.

3 PROCEDURE

If a beamline would like to report vibration issues related to construction at and around the APS, the following protocol is to be followed:

- 1) Beamline staff will contact the on-duty Floor Coordinator at 2-0101 to report suspected construction related vibration issues.
 - a) If the Floor Coordinator is not available, contact Bruce Glagola or Nena Moonier.
 - b) Glagola or Moonier will contact Vibration Task Force Subteam 4
- 2) The Floor Coordinator will alert a member of the Vibration Task Force Subteam 4, with contact in the following order:
 - a) Mark Beno (2-3507)
 - b) Julie Cross (2-0592)
 - c) Stefan Vogt (2-3071)
 - d) Dale Brewe (2-0582)
- 3) Subteam 4 member will contact the beamline to discuss the issue being observed and understand the issue.
- 4) Subteam 4 member will observe vibration data (electron beam motion, accelerometer readings, beam position monitors, or other).
- 5) Subteam 4 member will decide whether to recommend or to not recommend halting any activities at construction site.
- 6) Subteam 4 member will take the issue to the APS ALD Office, with contact in the following order:
 - a) Rod Gerig (2-5710)
 - b) Denny Mills (2-5680)
 - c) Brian Stephenson (2-7990)
- 7) The APS ALD Office will determine if the construction activity must be suspended.
- 8) If APCF construction activity is to be suspended, the ALD Office will contact
 - a) Steve Hunsberger (847-975-9428)

PROCEDURE	Page 5 of 5
Procedure #:	3.1.52
Revision #:	0

- b) Jesse Adams (224-532-3882) or
- c) Pat Hogan (312-656-1672)
- 9) If other construction activity is to be suspended, the ALD Office will contact
 - a) Rick Janik (2-6068) or
 - b) John Sidarous (2-6067)
- 10) If construction activity is to be suspended, the ALD Office will contact a member of Subteam 4, and Subteam 4 will notify the MCR. Information to be relayed is the type of construction activity that is stopped and if/when restart permission is given. The MCR will make an announcement if a construction activity is suspended.

4 FEEDBACK AND IMPROVEMENT

If you are using this procedure and have comments or suggested improvements for it, please go to the [APS Policies and Procedures Comment Form](#) * to submit your input to a Procedure Administrator. If you are reviewing this procedure in workflow, your input must be entered in the comment box when you approve or reject the procedure.

Instructions for execution-time modifications to a policy/procedure can be found in the following document: Field Modification of APS Policy/Procedure ([APS 1408152](#)).

* http://centraldocs.aps.anl.gov/comment_form.php